

Reproducibility  
10 ng Of Analyte on Column

Run #	Peak number			
	1	2	3	4
1	30.3	46.1	52.9	29.9
2	29.8	46.7	54.1	30.1
3	29.4	46.1	52.9	30.4
4	30.3	45.0	52.2	30.3
5	29.1	46.5	51.3	29.4
6	29.2	47.0	52.8	30.8
Mean	29.68	46.23	52.70	30.15
STDV	0.53	0.70	0.92	0.48
CV	0.018	0.015	0.018	0.016
<b>%RSD</b>	<b>1.80</b>	<b>1.51</b>	<b>1.75</b>	<b>1.58</b>

1=Fructose, 2=Glucose, 3=Lactose, 4=Sucrose

Reproducibility  
100 ng Of Analyte on Column

Run #	Peak number			
	1	2	3	4
1	294.5	478.0	557.0	332.2
2	295.1	479.2	553.2	333.6
3	300.5	479.2	548.9	331.0
4	294.6	476.2	555.5	333.6
5	298.3	465.4	553.3	327.3
6	299.1	475.0	557.5	325.5
Mean	297.0	475.5	554.2	330.5
STDV	2.6	5.2	3.2	3.4
CV	0.009	0.011	0.006	0.010
<b>%RSD</b>	<b>0.88</b>	<b>1.10</b>	<b>0.57</b>	<b>1.03</b>

1=Fructose, 2=Glucose, 3=Lactose, 4=Sucrose